

HAROLD JOHNSON

1826 WEST 169th STREET



Optical Laboratories, Inc.

GARDENA, CALIF. 90247 • PHONE: (310) 327-3051

FINAL INSPECTION

CUSTOMER *Caltech*PIECE NO. ☐SAMPLING ☐

CONTRACT NO.

DWG. NO. *9EDM-O-CA3 Camera 3*P. O. NO. *9139500*JOB NO. *1204045*

OPERATION	NO. OF PIECES	DIMENSIONS	DATE	SAMPLING	PIECE NO.	INSPECTOR
SURFACE QUALITY	1	<i>69/40</i>	<i>8/13/12</i>	<i>100%</i>		
SURFACE FLATNESS						
SURFACE SPHERICITY	1	<i>R1R2 3fr. x 1/2 fr. irr.</i>	<i>8/13/12</i>	<i>100%</i>		
THICKNESS	1	<i>.7874 ± .002 / .789</i>	<i>8/13/12</i>	<i>100%</i>		
DIAMETER	1	<i>2.4252⁺⁰ - .001 / 2.4246</i>	<i>8/13/12</i>	<i>100%</i>		
LENGTH						
WIDTH						
ANGLES						
FOCAL LENGTH						
CENTERING	1	<i>.0004</i>	<i>8/13/12</i>	<i>100%</i>		
T. I. R.						
RADIUS	1	<i>R1 2.5199 is 2.5201 R2 4.498 is 4.4972</i>	<i>8/13/12</i>	<i>100%</i>		
CHAMFER						
PARALLELISM						
REFLECTIVITY						
ADHERANCE						
CERTIFICATIONS:	1	<i>on file</i>	<i>8/13/12</i>	<i>100%</i>		
MATERIAL	1	<i>PBL26Y</i>	<i>8/13/12</i>	<i>100%</i>		
TRANSMISSION						

Disposition: _____

Mass 128.2 gm

Inspector _____