

HAROLD JOHNSON

1826 WEST 169th STREET



Optical Laboratories, Inc.

GARDENA, CALIF. 90247 • PHONE: (310) 327-3051

FINAL INSPECTION

CUSTOMER *Caltech*PIECE NO. ☐SAMPLING ☐

CONTRACT NO.

DWG. NO. *SEDM-0-CA5 Camera 5*P. O. NO. *3139500*JOB NO. *1204047*

OPERATION	NO. OF PIECES	DIMENSIONS		DATE	SAMPLING	PIECE NO.	INSPECTOR
SURFACE QUALITY	<i>1</i>	<i>60/40</i>	<i>OK</i>	<i>8/23/12</i>	<i>100%</i>		
SURFACE FLATNESS	<i>1</i>						
SURFACE SPHERICITY	<i>1</i>	<i>E1, R2 3fr. X 1/2 fr. irr.</i>	<i>OK</i>	<i>8/23/12</i>	<i>100%</i>		
THICKNESS	<i>1</i>	<i>.1575⁺.0021 .1557</i>		<i>8/23/12</i>	<i>100%</i>		
DIAMETER	<i>1</i>	<i>1.9685⁺.0011 1.9681</i>		<i>8/23/12</i>	<i>100%</i>		
LENGTH							
WIDTH							
ANGLES							
FOCAL LENGTH							
CENTERING	<i>1</i>	<i>.0004</i>	<i>OK</i>	<i>8/23/12</i>	<i>100%</i>		
T. I. R.							
RADIUS	<i>1</i>	<i>3.090 is 3.0901 8.170 is 8.1705</i>		<i>8/23/12</i>	<i>100%</i>		
CHAMFER							
PARALLELISM							
REFLECTIVITY							
ADHERANCE							
CERTIFICATIONS:	<i>1</i>	<i>on file</i>		<i>8/23/12</i>	<i>100%</i>		
MATERIAL	<i>1</i>	<i>PBM2Y</i>		<i>8/23/12</i>	<i>100%</i>		
TRANSMISSION							

Disposition: _____



Inspector _____