

HAROLD JOHNSON

1826 WEST 169th STREET



Optical Laboratories, Inc.

GARDENA, CALIF. 90247 • PHONE: (310) 327-3051

## FINAL INSPECTION

CUSTOMER *Caltech*PIECE NO. ☐SAMPLING ☐

CONTRACT NO.

DWG. NO. *GEDM-0-RC16 Rev. B*P. O. NO. *5139500*

JOB NO.

OPERATION	NO. OF PIECES	DIMENSIONS	DATE	SAMPLING	PIECE NO.	INSPECTOR
SURFACE QUALITY	<i>1</i>	<i>60/40</i>	<i>OK</i>	<i>9/13/12</i>	<i>100%</i>	
SURFACE FLATNESS	<i>1</i>					
SURFACE SPHERICITY	<i>1</i>	<i>2.22 3 fr. x 1/2 fr. i.e.e</i>	<i>OK</i>	<i>9/13/12</i>	<i>100%</i>	
THICKNESS	<i>1</i>	<i>.1181 ± .0039</i>		<i>9/13/12</i>	<i>100%</i>	
DIAMETER	<i>1</i>	<i>1.1811<sup>+00</sup> - .001</i>   <i>1.1808</i>		<i>9/13/12</i>	<i>100%</i>	
LENGTH						
WIDTH						
ANGLES						
FOCAL LENGTH						
CENTERING	<i>1</i>	<i>.0019</i>	<i>OK</i>	<i>9/13/12</i>	<i>100%</i>	
T. I. R.						
RADIUS	<i>1</i>	<i>51 1.1211 is 1.121</i>   <i>52 2.216 is 2.2149</i>		<i>9/13/12</i>	<i>100%</i>	
CHAMFER	<i>1</i>	<i>.0196 x 45°</i>	<i>OK</i>	<i>9/13/12</i>	<i>100%</i>	
PARALLELISM						
REFLECTIVITY						
ADHERANCE						
CERTIFICATIONS:	<i>1</i>	<i>on file</i>		<i>9/13/12</i>	<i>100%</i>	
MATERIAL	<i>1</i>	<i>PBL6Y (furn.)</i>		<i>9/13/12</i>	<i>100%</i>	
TRANSMISSION						

Disposition: \_\_\_\_\_



Inspector \_\_\_\_\_